


BLANK PAGE





Indian Standard





SPECIFICATION FOR

"KE AFRIRMED

FORCEPS, EXTRACTION, DENTAL, LOWER INCISORS, CUSPIDS AND BICUSPIDS

- 1. Scope Dimensional and other requirements for dental forceps for the extraction of lower incisors, cuspids and bicuspids.
- 2. Shape and Dimensions As shown in Fig. 1 and 2.





All dimensions in millimetres.


FIG. 1 FORCEPS, EXTRACTION, DENTAL, LOWER INCISORS, CUSPIDS AND BIGUSPIDS

- 3. Weight Shall be 145 g.
- 4. Other Requirements All the requirements given in IS: 6868-1972 General requirements for forceps, extraction, dental' shall apply.
- 4.1 Performance Load Shall be 29 kg.

Adopted 28 December 1972

@ May 1973, ISI

Gr 1



All dimensions in millimetres.

FIG. 2 DETAILS OF WORKING-ENDS, LOWER INCISORS, CUSPIDS AND BICUSPIDS